



# Surface T/S Data RV "Heincke" HE437

**Data Processing Report** 

## Contents

1	Introduction	1
2	Workflow	1
3	Cruise details	2
4	Sensor	2
5	Processing Report	3

Contact: Gerd Rohardt Alfred-Wegener-Institute Am Handelshafen 12, D-27570 Bremerhaven, GERMANY Mail: info@awi.de

Processing Agency: FIELAX Schleusenstr. 14, D-27568 Bremerhaven, GERMANY Mail: info@fielax.de

Ref.: HE437_TSG.pdf	Vers.: 1	Date: 2016/04/11	Status: final	]
---------------------	----------	------------------	---------------	---



### **1** Introduction

This report describes the processing of raw data acquired by the thermosalinograph on board RV "Heincke" during expedition HE437 to receive cleaned up and drift corrected salinity data.

### 2 Workflow

The different steps of processing are visualized in Figure 2. Unvalidated data of sensor, internal and external temperature are extracted from the DAVIS SHIP data base (https://dship.awi.de) in a 1-second interval for cruises from 2009 to 2014. The Salinity was calculated by applying the Practical Salinity Scale 1978 (PSS-78). Furthermore the sound velocity was derived by using the Del Grosso equation.

As first step, a basic cleanup was performed to remove missing or flagged data. Since the salinity measurements in coastal areas (e.g. rivers and ports) are less reliable, measurements in a buffer of 2 nautical miles (NM)along the coast are filtered. In the norwegian area (fjords) the buffer is set to 200 meters (0.108 NM). After the exclusion of data outside the speed interval of 0.5 kn to 15 kn, the salinity is driftcorrected with lab calibration data. In the next processing step the difference between the external and internal temperature is taken to identify an unproper usage of the thermosalinograph. This filter is ignored if more than 90% of the data would get removed. After despiking, a visual screening is performed to enhance the data quality. In the last step the temporal resolution is reduced to 5-minutes-means.




Figure 1: Workflow of TSG data processing



## 3 Cruise details

Vessel name	RV "Heincke"
Cruise name	HE437
Cruise start	24.11.2014 Bremerhaven
Cruise end	29.11.2014 Bremerhaven
Cruise duration	5 days

### 4 Sensor

Thermosalinograph:Seabird SEACAT SBE21 (SN: 3334)External Temperature:SBE38




Figure 2: Cruisemap of HE437.



## **5 Processing Report**

#### **Database Extraction**

Data source	DSHIP database (dship.awi.de)
Exported values	518401
First dataset	2014-11-24T00:00:03 UTC
Last dataset	2014-11-30T00:00:00 UTC

#### **Automatic Validation**

The following thresholds were applied for the automatic flagging of the position data:

Min. speed	Minimum 0.5 kn between two datapoints.	
Max. speed	d Maximum 15 kn between two datapoints.	
GeoBuffer	0.1080 NM around Norway, 2 NM anywhere else	
Temperature Maximum T-difference of 5 K.		

### Flagging result


Filter	Data left (abs.)	Data left (rel.)	Data removed (abs.)	Data removed (rel.)
Raw data	518401	100 %	—	—
Basic	273893	52.83%	244508	47.17%
Geo	232247	44.80 %	286154	55.20 %
Speed	225332	43.47 %	293069	56.53 %
Temperature	225332	43.47 %	293069	56.53 %
Despike	219791	42.40 %	298610	57.60 %
Manual	217262	41.91 %	301139	58.09%
5-min-Mean	780	0.15%	517621	99.85 %


#### Sensordrift

Last calibration	29.01.2013
Current calibration	29.04.2015
Start of deployment	09.12.2013
End of deployment	19.02.2015
Scaled drift	-9.3822e-004 [PSU/month]
Minimal offset	1.0840e-002 [PSU]
Maximal offset	1.0938e-002 [PSU]

# 

#### **Process evolution**








Figure 4: Salinity after basic filter.






Figure 5: Salinity after geofilter.

# 








Figure 7: Salinity after temperature filter.




Figure 8: Salinity after despike.

# 




Figure 9: Salinity after manual filter.




Figure 10: Salinity in 5-min-mean values.



#### **Result file**

Text File (HE437\_surf\_oce.tab):

The format is a plain text (tab-delimited values) file.

Column separator	Tabulator "\t"
Column 1	Date and time expressed according to ISO 8601
Column 3	Latitude in decimal format, unit degree
Column 4	Longitude in decimal format, unit degree
Column 5	Depth below water surface, unit meter
Column 6	Temperature, unit degree
Column 7	Salinity, unit PSU

Processing Report (HE437\_TSG.pdf): This PDF document.